

York Region Infection Prevention And Control Lapse Report

Initial Report					
Premise/Facility under investigation (name and address Armadale Family and Cosmetic Dentistry 7720 Markham Road Unit A4 Markham, Ontario L3S 3K1	ess)				
Type of Premises/Facility Dental Office					
Date Board of Health became aware of IPAC lapse (yyyy/mm/dd) 2023/01/13		Date of Initial Report posting (yyyy/mm/dd) 2023/01/19			
Date of Initial Report update(s) (if applicable) (yyyy/mm/dd) 2023/01/26			How the IPAC lapse was identified Complaint		
Summary Description of the IPAC Lapse As a result of a complaint, York Region Public Health co Region Public Health noted concerns with handling of m per best practices and manufacturer's instructions for us	nulti-use de				
IPAC Lapse Investigation	Yes	No	N/A	Please provide further details/steps	
Did the IPAC lapse involve a member of a regulatory college?	\boxtimes				
If yes, was the issue referred to the regulatory college?				Royal College of Dental Surgeons of Ontario College of Dental Hygienists of Ontario	
Were any corrective measures recommended and/or implemented?					
Please provide further details/steps	 Corrective measures: Operator to ensure that for all dental equipment/devices requiring sterilization, a sterilization process used is in accordance with the "PIDAC Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings, 3rd Edition, May 2013". Operator to ensure all staff responsible for reprocessing of dental/medical equipment/devices follow the established policies and procedures, which must be in accordance with the "PIDAC Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings, 3rd Edition, May 2013" and the Manufacturer's Instructions for Use (MIFU) for each equipment/device. Operator to cease use of newly purchased non-sterile critical and semi-critical dental/medical equipment/devices until they are inspected and reprocessed prior to use, according to their intended use, as per the MIFU. Operator to cease re-using a single-use sterile syringe when accessing dental solutions intended for multi-use (e.g., irrigants, filing and chelating agents) and to ensure that a used syringe is never reattached to a multi-dose container/cartridge, etc. Operator to ensure all sterile single-use critical and semi-critical equipment/devices are discarded where tamper seals are broken, and/or sterility of the equipment/device is compromised. 				
Date any order(s) or directive(s) were issued to the c 2023/01/11		erator (if	applica	able) (yyyy/mm/dd)	
Initial Report Comments and Contact Infor Any additional Comments: (Please do not include an		al inforn	nation o	or personal health information)	



1-800-361-5653

York Region Infection Prevention And Control Lapse Report

	iniconditi revention And Control Eapse Report
Health Connection	
Telephone Number	Email Address
1-800-361-5653	Health.inspectors@york.ca
Final Report	
Date of Final Report posting (yyyy/mm/d 2023/01/26	d)
Date any order(s) or directive(s) were iss	ued to the owner/operator (if applicable) (yyyy/mm/dd)
Brief description of corrective measures All corrective measures listed above have b	
Date of all corrective measures were con 2023/01/24	firmed to have been completed (yyyy/mm/dd)
Final Report Comments and Cont	act Information
,	ot include any personal information or personal health information) 24, 2023. No concerns were noted at the time of re-inspection. a standing order.
If you have any further questions, please co Health Connection	ntact
Telephone Number	Email Address

Health.inspectors@york.ca