

Initial Report

Premises/Facility under investigation (name and address)

Ontario Diagnostic Centre
8077 Islington Avenue
Unit: 110
Woodbridge, Ontario L4L 7X7

Type of Premises/Facility

Clinical – Diagnostic

Date Board of Health became aware of IPAC lapse (yyyy/mm/dd) 2024- 04-03	Date of Initial Report posting (yyyy/mm/dd) 2024/05/10
--	--

Date of Initial Report update(s) (if applicable) (yyyy/mm/dd)	How the IPAC lapse was identified Complaint
--	---

Summary Description of the IPAC Lapse

- Cleaning and disinfection of reusable medical equipment/devices was not conducted in accordance with “Public Health Ontario: Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices In All Health Care Settings, 3rd Edition, May 2013”.
- There were disinfectants, used for reprocessing of reusable equipment, that did not bear a Health Canada Drug Identification Number, Natural Product Number or Class 2 Device License.

IPAC Lapse Investigation	Yes	No	N/A	Please provide further details/steps
Did the IPAC lapse involve a member of a regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
If yes, was the issue referred to the regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were any corrective measures recommended and/or implemented?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

Please provide further details/steps	<p>Corrective measures for Premises/Facility:</p> <ul style="list-style-type: none"> • Reprocess (clean, and disinfect or sterilize) re-usable medical equipment/devices after each use in accordance with the “Public Health Ontario: Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices In All Health Care Settings, 3rd Edition, May 2013”. • Use disinfectants for the reprocessing of reusable equipment/devices that have an expiry date, a Drug Identification Number (DIN), Natural Product Number (NPN), and/or Medical Device License (MDAL) with Health Canada (with exception of Chlorine Bleach). • Provide personal protective equipment (PPE), such as gowns, gloves, mask, and eye protection in reprocessing area, and train staff on routine practices including conducting risk assessment, hand hygiene and use of PPE. • Evaluate all staff who perform reprocessing to ensure they have demonstrated competency in the
--------------------------------------	---



Infection Prevention and Control Lapse Report

reprocessing process of reusable medical equipment/devices.

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

Order Issued 2024/03/01.

Initial Report Comments:

An Order was issued on March 1, 2024, ordering operator to implement all corrective measures related to reprocessing of re-usable medical equipment/devices.

Any additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact
Health Connection

Telephone Number
1-800-361-5653

Email Address
Health.inspectors@york.ca

Final Report

Date of Final Report posting (yyyy/mm/dd)

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

Brief description of corrective measures taken

Date of all corrective measures were confirmed to have been completed (yyyy/mm/dd)

Reinspection conducted and all corrective measure were confirmed to have been completed 2024/02/27

Final Report Comments and Contact Information

Any Additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact
Health Connection

Telephone Number
1-800-361-5653

Email Address
Health.inspectors@york.ca