

Initial Report

Premises/Facility under investigation (name and address)

Woodbridge Square Dental Clinic
7600 Weston Road, Unit 49
Woodbridge, ON., L4L 8B7

Type of Premises/Facility

Clinical - Dental

Date Board of Health became aware of IPAC lapse (yyyy/mm/dd)	Date of Initial Report posting (yyyy/mm/dd)
2025/11/21	2025/12/03

Date of Initial Report update(s) (if applicable) (yyyy/mm/dd)	How the IPAC lapse was identified
	Referral

Summary Description of the IPAC Lapse

- Reusable semi-critical and critical dental instruments/devices were not sterilized between patients as required and in accordance with the “PIDAC Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings, 3rd Edition, May 2013”.
- Dental instruments/devices that have been reprocessed were not clearly distinguished from those that had not been reprocessed, as required by the best practice standards in Infection Prevention and Control.
- Sterilized dental instruments/devices were not verified for chemical indicator (CI) colour change before use on patients, as required by the best practice standards in Infection Prevention and Control.

IPAC Lapse Investigation	Yes	No	N/A	Please provide further details/steps
Did the IPAC lapse involve a member of a regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Royal College of Dental Surgeons of Ontario (RCDSO) College of Dental Hygienists of Ontario (CDHO)
If yes, was the issue referred to the regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were any corrective measures recommended and/or implemented?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

Please provide further details/steps

- Corrective measures for Premises/Facility:**
- Monitor and verify sterilization verification processes, including physical, chemical, and biological indicators, prior to releasing dental instruments/devices for use on patients, in accordance with the “PIDAC Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings, 3rd Edition, May 2013”.
 - Check and verify chemical indicators (CIs) for appropriate colour change before instruments/devices are used. Where a failed CI is found, reprocess the instruments/devices.
 - Implement a process to clearly differentiate dental instruments/devices that have been reprocessed from those that have not been reprocessed.



Infection Prevention and Control Lapse Report

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

HPPA Section 13 Order issued on 2025/11/07.

Initial Report Comments:

HPPA Section 13 Order was issued on 2025/11/07, requiring the operator to correct procedures related to monitoring and verification of sterilization processes, including monitoring and verification of physical, chemical, and biological indicators, and to implement a system that clearly differentiates reprocessed instruments from those that have not been reprocessed.

Any additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact:

Health Connection

Telephone Number

1-800-361-5653

Email Address

Health.inspectors@york.ca

Final Report

Date of Final Report posting (yyyy/mm/dd)

2025/12/03

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

HPPA Section 13 Order issued on 2025/11/07.

Brief description of corrective measures taken

Date of all corrective measures were confirmed to have been completed (yyyy/mm/dd)

2025/11/07

Final Report Comments and Contact Information

Any Additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact:

Health Connection

Telephone Number

1-800-361-5653

Email Address

Health.inspectors@york.ca